

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com
Email USA: techserv@sial.com
Outside USA: eurtechserv@sial.com

## **Product Specification**

**Specification** 

None Detected

Product Name:

**TEST** 

RNAse Detection

Acrylamide/Bis-acrylamide - BioReagent, for molecular biology, 29:1 (ratio)

Product Number: A2792

MDL: MFCD00080848

 $H_2C \underset{\mathsf{NH}_2}{\overset{\circ}{\bigcup}} H_2C \underset{\mathsf{N}}{\overset{\circ}{\bigcup}} \underset{\mathsf{N}}{\overset{\circ}{\bigcup}} CH_2$ 

Appearance (Color)	White
Appearance (Form)	Powder
Water insolubles	≤ 0.005 %
Acrylic Acid	< 0.005 %
Conductivity	<ul><li>_</li><li>10</li></ul>
micromho, 40% solution	
Lead (Pb)	≤ 5 ppm
Magnesium (Mg)	≤ 3 ppm
Iron (Fe)	≤ 2 ppm
Protease Detection	None Detected
Ratio	26.8 - 31.2
26.8:1 to 31.2:1 acrylamide / bis-acrylamide ratio	
DNAse, Exonuclease Detection	None Detected
NICKase, Endonuclease Detection	None Detected

Specification: PRD.2.ZQ5.10000009887.000

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.