

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.comEmail USA: techserv@sial.comOutside USA: eurtechserv@sial.com

Product Specification

Product Name:

Prestained Molecular Weight Marker – mol wt 26,600–180,000 Da

Product Number:

SDS7B2

Storage Temperature:

2 - 8 °C

TEST	Specification
Appearance (Color)	Blue
Appearance (Form)	Powder
SDS Electrophoresis	Pass
S8445 is used as a standard to determine the MW of 7 major bands on a 4-20% Laemmli gel.	
Western Blot	Pass
Seven major blue bands detected upon transfer to nitrocellulose membrane.	
Human Source Virus Testing	Conforms
Source material tested and found negative for antibodies to HIV, antibodies to HCV, and for HBsAg.	

Specification: PRD.0.ZQ5.10000059173

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.